

Certificate of Calibration

Customer: TEST COMPANY
100 PRODUCTION LANE
ANYWHERE, IL, 61109
111-111-1111

P.O. Number:

ID Number: 123456



Description: DIGITAL MULTIMETER, 6 1/2 DIGIT
Manufacturer: FLUKE
Model Number: 8845A
Serial Number: 123456
Technician: SCOTT CHAMBERLAIN

Calibration Date: 05/20/2024
Calibration Due: 05/31/2025
Procedure: METCAL FLUKE 8845A
Rev: 5/1/2020
Temperature: 71 °F
Humidity: 50 % RH

As Found Condition: IN TOLERANCE

Calibration Results: IN TOLERANCE

On-Site Calibration:

Comments: Sample Certificate of Calibration

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.

Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement unless otherwise noted.

TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration complies with all the requirements of ANSI/NCSL Z540-1-1994 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc.

Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.

S. COURNOYER, BRANCH MANAGER

Scott Chamberlain, QUALITY MANAGER

Calibration Standards

Asset Number	Manufacturer	Model Number	Date Calibrated	Cal Due
MELB-59	FLUKE	5730A/03	4/30/2024	4/30/2025
MELB-59A	FLUKE	5725A	2/12/2024	2/12/2025



Technical Maintenance, Inc.

3060 VENTURE LANE, SUITE 106, MELBOURNE, FL 32934

ANSI/NCSL Z540-1-1994